



ClinChek® - Control

Plasma Control lyophilised / Kontrollplasma lyophilisiert

FOR BIOGENIC AMINES / FÜR BIOGENE AMINE

Intended use:

ClinChek® Plasma Controls are used for internal quality assurance in clinical-chemical laboratories. These lyophilised controls are based on human plasma and are available with mean values in the normal as well as in the pathological range. After reconstitution the controls have to be prepared like patient samples in one series of analyses.

Reconstitution:

Add exactly 5.0 ml of HPLC-water to the vial and mix for 15 min. When all material is dissolved, the solution is ready to use.

Storage and stability:

This product will be stable until the expiration date when stored unopened at 2 - 8 °C.

After reconstitution the stability of the analytes is:

- 5 hours at 15 - 30 °C
 - 1 day at 2 - 8 °C
 - 10 days below -18 °C
(3 months for Serotonin)
- (avoid repeated freezing and thawing)

Notes:

The concentrations of the analytes are chosen in ranges where valid results can be obtained. According to quality assurance all ClinChek® Controls have to pass strict quality control procedures during manufacturing. RECIPE guarantees the same stability and constitution for each vial of one lot. The variation of the filling volume (CV) is < 1 %. The average residual moisture of this lot is 1.21 %.

Mean values:

The mean values and confidence intervals have been established at RECIPE, according to the Guideline of the German Medical Association on Quality Assurance (Rili-BAEK), with statistical methods.

Pack size:

ClinChek® Plasma Control
Level I
10 x 5 ml, **order no.: 8810**
Level II
10 x 5 ml, **order no.: 8811**
Level I, II
2 x 5 x 5 ml, **order no.: 8809**

Precautions:

The human plasma which was used for manufacturing the controls was tested for the following infectious markers and found negative: HIV1/2- and HCV-antibodies, Hepatitis B-surface antigen. Nevertheless, the plasma controls should be considered as potentially infectious and treated with appropriate care.

Zweckbestimmung:

ClinChek® Kontrollplasmen dienen der internen Qualitätssicherung im klinisch-chemischen Laboratorium. Es handelt sich um lyophilisierte Kontrollen auf Humanplasmabasis mit Sollwerten im normalen und pathologischen Bereich. Nach Rekonstitution werden die Kontrollproben analog zu den Patientenproben in einer Analysenserie aufgearbeitet.

Rekonstitution:

Zum Inhalt eines Fläschchens werden exakt 5.0 ml HPLC-Wasser gegeben und die Lösung unter gelegentlichem Umschwenken für ca. 15 Minuten stehen gelassen. Nach erneuter sorgfältiger Durchmischung kann die Lösung verwendet werden.

Lagerung und Haltbarkeit:

Dieses Produkt ist bis zum angegebenen Haltbarkeitsdatum stabil, wenn es ungeöffnet bei 2 - 8 °C gelagert wird.

Die Haltbarkeit der Analyten in der rekonstituierten Lösung beträgt:

- 5 Stunden bei 15 - 30 °C
- 1 Tag bei 2 - 8 °C
- 10 Tage bei < -18 °C
(3 Monate für Serotonin)
(nur einmal auftauen)

Anmerkungen:

Die Analytkonzentrationen liegen im gut messbaren Bereich. Im Rahmen der Qualitätssicherung werden alle ClinChek® Kontrollen bei der Herstellung einer strengen Qualitätsprüfung unterzogen. RECIPE garantiert für jedes Fläschchen einer Charge gleiche Haltbarkeit und Zusammensetzung. Die Abfüllpräzision (VK) ist < 1 %. Die durchschnittliche Restfeuchte dieser Charge beträgt 1.21 %.

Sollwerte:

Die Sollwerte und Vertrauensbereiche wurden von RECIPE entsprechend der Richtlinie der Deutschen Bundesärztekammer zur Qualitätssicherung (Rili-BÄK) mit statistischen Methoden ermittelt.

Packungsgröße:

ClinChek® Kontrollplasma
Level I
10 x 5 ml, **Best.-Nr.: 8810**
Level II
10 x 5 ml, **Best.-Nr.: 8811**
Level I, II
2 x 5 x 5 ml, **Best.-Nr.: 8809**

Vorsichtsmaßnahmen:

Das zur Herstellung der Kontrollen verwendete Humanplasma wurde auf folgende Infektionsmarker untersucht und für negativ befunden: HIV1/2- und HCV-Antikörper, Hepatitis B-Oberflächenantigen. Unabhängig davon sollten alle verwendeten Kontrollplasmen als potentiell infektiös angesehen und mit angemessener Sorgfalt behandelt werden.

ClinChek® - Control

Plasma Control, Level I

Kontrollplasma, Level I

REF	8810
LOT	1029
✉	2022-01

Analyte / Analyt	Method of Analysis / Analysenmethode	Unit / Einheit	Mean Value / Sollwert	Control Range / Kontrollbereich
Norepinephrine / Noradrenalin	RECIPE-HPLC	ng/l	347	277 - 416
		pmol/l	2048	1639 - 2458
Epinephrine / Adrenalin	RECIPE-HPLC	ng/l	102	81.2 - 122
		pmol/l	554	443 - 665
Dopamine / Dopamin	RECIPE-HPLC	ng/l	92.9	65.1 - 121
		pmol/l	607	425 - 789
Serotonin	RECIPE-HPLC	µg/l	114	91.1 - 137
		nmol/l	645	516 - 774

ClinChek® - Control

Plasma Control, Level II

Kontrollplasma, Level II

REF	8811
LOT	1029
✉	2022-01

Analyte / Analyt	Method of Analysis / Analysenmethode	Unit / Einheit	Mean Value / Sollwert	Control Range / Kontrollbereich
Norepinephrine / Noradrenalin	RECIPE-HPLC	ng/l	2222	1778 - 2667
		pmol/l	13134	10508 - 15761
Epinephrine / Adrenalin	RECIPE-HPLC	ng/l	594	475 - 713
		pmol/l	3243	2594 - 3891
Dopamine / Dopamin	RECIPE-HPLC	ng/l	563	423 - 704
		pmol/l	3678	2759 - 4598
Serotonin	RECIPE-HPLC	µg/l	341	273 - 410
		nmol/l	1932	1546 - 2319